



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/798,601

03/10/2004

Edward I. Wulfman

89000.3012NP

6339

20601 7590 04/28/2008
SPECKMAN LAW GROUP PLLC
1201 THIRD AVENUE, SUITE 330
SEATTLE, WA 98101

EXAMINER

MENDOZA, MICHAEL G

ART UNIT

PAPER NUMBER

3734

MAIL DATE

DELIVERY MODE

04/28/2008

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/798,601	Applicant(s) WULFMAN, EDWARD I.	
	Examiner MICHAEL G. MENDOZA	Art Unit 3734	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 10 August 2007.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-21 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-21 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date <u>9/2/05, 12/18/06, 8/22/07</u> . | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Rejections - 35 USC § 112

1. The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

2. Claims 18-21 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.
3. Claim 18 recites the limitation "the tubular structure" in line 4. There is insufficient antecedent basis for this limitation in the claim.
4. Claim 20 recites the limitation "drive system" in line 2. There is insufficient antecedent basis for this limitation in the claim.

Claim Rejections - 35 USC § 102

5. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:


A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

6. Claims 1-3, 5-9, 11-16, 18, and 19 are rejected under 35 U.S.C. 102(b) as being anticipated by Samson et al. 6143013.
7. Samson et al. teaches a medical tubular structure comprising: a lumen; an overlying layer; a support layer having a contiguous coil element, a braid element or a weave element including a plurality of loops, the support layer attached to the overlying layer at a bonding point and further a free portion, wherein the free portion is slippable

Art Unit: 3731

relative to the overlying layer (only part that is bonded is shown of 406, the remaining is free); a thermally shrinkable sheath having a plurality of etches on at least its interior surface (col. 15, lines 34-36); wherein the sheath encases the support layer by heat-reduction of 25 percent or less; wherein the sheath comprises PTFE, Teflon, FEP and/or PFA (col. 11, lines 25-34); wherein the support layer includes the contiguous coil element comprised of a wire and a plurality of gaps between each loop; wherein the length of each gap is about 10-200 percent of the width of the wire (see figures); and an operating head (most distal section).




8. Claims 18, 20 and 21 are rejected under 35 U.S.C. 102(b) as being anticipated by Ressemann et al. 5501694.

9. teaches an intracorporeal medical device comprising: an operating head, and a catheter comprising: a lumen; a support layer encasing the lumen and having a contiguous coil element, a braid element or a weave element including a plurality of

Art Unit: 3731

loops, the support layer attached to the overlying layer at a bonding point and further a free portion, wherein the free portion is slippable relative to the overlying layer; a thermally shrinkable sheath (col. 12, lines 6-8) having a plurality of etches on at least its interior surface; wherein the sheath encases the support layer by heat-reduction of 25 percent or less; a drive shaft; and a control system.



Claim Rejections - 35 USC § 103

10. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

Art Unit: 3731

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

11. Claims 4, 10, and 17 are rejected under 35 U.S.C. 103(a) as being unpatentable over Samson et al.

12. Samson et al. teaches a device that is flexible yet highly resistant to kinking. Samson et al. fails to teach about what radius the device can go without kinking. It would have been obvious to one having ordinary skill in the art at the time the invention was made to adjust the flexibility of the device to a desired range before kinking, since it has been held that where the general condition of a claim are disclosed in the prior art, discovering the optimum or workable ranges involves only routine skill in the art. *In re Aller*, 105 USPQ 233.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to MICHAEL G. MENDOZA whose telephone number is (571)272-4698. The examiner can normally be reached on Mon.-Fri. 9:00 a.m. - 5:00 p.m..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Todd Manahan can be reached on (571) 272-4713. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3731

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/M. G. M./

Examiner, Art Unit 3734

/Todd E Manahan/

Supervisory Patent Examiner, Art Unit 3731